

Technical Data Sheet

12/07/2017

**DMEM Low Glucose w/o L-Glutamine w/ Sodium Pyruvate**

Product code :  
LM-D1100

<b><u>Theoretical pH:</u></b>	7.3 ± 0.3
<b><u>Osmolality:</u></b>	320 mOsm/kg ±10%
<b><u>Colour:</u></b>	Red, clear solution
<b><u>Storage conditions:</u></b>	+2 to 8°C
<b><u>Shelf life:</u></b>	24 months
<b><u>Endotoxin:</u></b>	<1 EU/ml
<b><u>Composition:</u></b>	Available on request
<b><u>Sterility Tests</u></b>	Bacteria aerobic-anaerobic Bacteria, strictly anaerobic Fungi/Yeast

**Cell Growth Test**

Medium tested for the ability to support Hela cell growth.

**Recommended Use:**

- Respect storage conditions of the product
- Do not use the product after its expiry date
- Store product in an area protected from light (not necessary for saline solutions).
- Manipulate the product in aseptic conditions (e.g. : under laminar air flow)
- Wear clothes adapted to the manipulation of the product to avoid contamination (e.g. : gloves, mask, hygiene cap, overall...)

The product is intended to be used in vitro, in laboratory only. Do not use it in therapy, human or veterinary applications.

## **Description:**

Lots of modifications of Eagle's medium have been developed since the creation of the first formulation. The most used Eagle's medium is the Dulbecco's Modified Eagle's Medium (DMEM).

It is a modification of Basal Medium Eagle (BME) that contains an increased concentration of amino acids and vitamins and also supplementary components. DMEM Low Glucose corresponds to the original formulation with 1000mg/L of glucose; it was used to culture embryonic mouse cells.

## **Uses:**

Supplements such as antibiotics should be added as sterile supplements to the medium. Storage conditions and shelf life of supplemented product will be affected by the nature of the supplements. Sterile serum should not be re-filtered before or after being added to sterile medium because growth-promoting capacity may be reduced upon re-filtration. Add 20ml/L of L-Glutamine 200mM before using this medium.

## **Indication of Deterioration:**

Medium should be clear and free of flocculent material.

Do not use if the medium is cloudy or contains precipitate.

Other evidence of deterioration may include colour change or degradation of physical or performance characteristics.